Color All-in-one Reverse Transcription Kit (with DNase)

Catalog No.: RT3CP

Description

The **EZBioscience**® Color All-in-one Reverse Transcription Kit (with DNase) is a new generation of rapid reverse transcription kit for genomic DNA removal and reverse transcription reaction in one step. Compared with the EZBioscience® Color Reverse Transcription Kit (with gDNA Remover) (Cat. No.: A0010CG), this kit does not require a separate genomic DNA removal reaction, which is more convenient and can effectively reduce the risk of contamination and RNA degradation caused by complex sampling process. The kit mainly includes five tubes of reagents: Enzyme Mix, 5× Color All-in-one RT Buffer, NC Buffer, Oligo dT18 (20x) and Random Hexamer (20x). Among them. The Enzyme Mix contains DNase. RNase Inhibitor and reverse transcriptase; The 5× Color All-in-one RT Buffer contains buffer, dNTPs, blue dye, etc.; Oligo dT18 primer and Random Hexamer are packaged separately, which is convenient for adding gene-specific primers as needed.

Blue dye in the 5× Color All-in-one RT Buffer could avoid pipetting errors. It can be used with the **EZBioscience®** Color SYBR Green qPCR Mix (Cat. No.: A0012, A0012-R1, A0012-R2), which is supplemented with an inert red dye. Mixing cDNAs, primers and other components with the **EZBioscience®** Color SYBR Green qPCR Mix together in a qPCR reaction turns the solution into purple, which provides a visual aid when pipetting and decreases the risk of pipetting errors during reaction setup, especially when using white reaction vessels. And the dyes do not affect the specificity or sensitivity of qPCR reactions. So, these two kinds of kits are recommended to use together to get optimal results.

The reverse transcriptase in the Enzyme Mix is a genetic engineered enzyme based on M-MLV (RNase H-) reverse transcriptase. And the multiple site-mutations of reverse transcriptase can obviously increase its affinity to RNA templates and its strand extending ability, which make reverse transcription reaction more efficient. Moreover, this transcriptase is rather resistant to common reverse transcriptase inhibitors. At the same time, the kit uses the latest optimized reaction system to further improve the reverse transcription efficiency. The reverse transcription reaction products are recommended for PCR, qPCR and gene cloning.

Components

Components	RT3CP	RT3CP-L	
Componente	(100 Rxns)	(500 Rxns)	
Enzyme Mix	200 µl	200 μ l × 5 tubes	
5× Color All-in-one		400 5	
RT Buffer	400 µl	400 μl × 5 tubes	
NC Buffer	20 µl	20 μl × 5 tubes	
Oligo dT18 (20×)	110 µl	110 μ l × 5 tubes	
Random Hexamer	440	110 μ l × 5 tubes	
(20×)	110 µl		
ddH₂O (Nuclease⁻)	1 ml	1 ml × 5 tubes	

Storage

Store at -20°C.

Caution

Avoid RNase contamination

Please keep the environment of experiment clean. Clean gloves and mask should be worn during the experiment. Centrifuge tubes, tips and other supplies used in the experiment must be RNase-free.

Protocol

Thaw template RNA on ice. Thaw Enzyme Mix, 5× Color All-in-one RT Buffer, NC Buffer, ddH₂O (Nuclease⁻), Oligo dT18 (20×), Random Hexamer (20×) or gene-specific primer at room temperature (15 ~ 25°C) or on ice.

Mix each solution by inverting the tubes up and down 10 times. Centrifuge briefly to collect residual liquid from the sides of the tubes, and then put on ice for later use.

1. Set up reactions

1A. If used for reverse transcription reaction of mRNA/ circRNA/ lncRNA:

(1) Prepare the one-step reaction for genomic DNA elimination and reverse transcription on ice according to the following table.

Add 1 μ g total RNA (10 pg ~ 2 μ g adjustable) or 200 ng poly(A) mRNA (10 pg ~ 500 ng adjustable) to a sterile, Nuclease-free tube, then add 2 μ l of Enzyme Mix, 4 μ l of 5× Color All-in-one RT Buffer, 1 μ l of Oligo dT18 (20×), 1 μ l of Random Hexamer (20×), and then add ddH₂O (Nuclease) to up to 20 μ l:

Components	Volume (20 μl)
total RNA	1 µg
or poly(A) mRNA	200 ng
Enzyme Mix	2 µl
5× Color All-in-one RT Buffer	4 µl
Oligo dT18 (20×)	1 µl
Random Hexamer (20×)	1 µl
ddH ₂ O (Nuclease ⁻)	up to 20 μl

Note: When performing circRNA reverse transcription, Oligo dT18 (20×) can be used or not.

(2) (Optional) Prepare No Reverse-Transcriptase Control reaction.

No Reverse-Transcriptase Control is a negative control reaction without reverse transcriptase, which is used to detect whether there is genomic DNA residue in RNA template.

Add 1 μ g total RNA (10 pg ~ 2 μ g adjustable) or 200 ng poly(A) mRNA (10 pg ~ 500 ng adjustable) to a sterile, Nuclease-free tube, then add 4 μ l of 5× Color All-in-one RT Buffer, 2 μ l of NC Buffer, 1 μ l of Oligo dT18 (20×), 1 μ l of Random Hexamer (20×), and then add ddH₂O (Nuclease⁻) to up to 20 μ l:

Components	Volume (20 μl)
total RNA	1 μg
or poly(A) mRNA	200 ng
5× Color All-in-one RT Buffer	4 µl
NC Buffer	2 µl
Oligo dT18 (20×)	1 µl
Random Hexamer (20×)	1 µl
ddH ₂ O (Nuclease ⁻)	up to 20 μl

- 1B. If used for reverse transcription reaction of target specific gene:
- (1) Prepare the one-step reaction for genomic DNA elimination and reverse transcription on ice according to the following table.

Add 1 μ g total RNA (10 pg ~ 2 μ g adjustable) or 200 ng poly(A) mRNA (10 pg ~ 500 ng adjustable) to a sterile, Nuclease-free tube, then add 2 μ l of Enzyme Mix, 4 μ l of 5× Color All-in-one RT Buffer, 1 μ l of gene-specific primer (2 μ M), and then add ddH₂O (Nuclease⁻) to up to 20 μ l:

Components	Volume (20 µl)
total RNA	1 µg
or poly(A) mRNA	200 ng
Enzyme Mix	2 µl
5× Color All-in-one RT Buffer	4 µl
gene-specific primer (2 µM)	1 µl
ddH ₂ O (Nuclease ⁻)	up to 20 μl

(2) (Optional) Prepare No Reverse-Transcriptase Control reaction.

Add 1 μ g total RNA (10 pg ~ 2 μ g adjustable) or 200 ng poly(A) mRNA (10 pg ~ 500 ng adjustable) to a sterile, Nuclease-free tube, then add 4 μ l of 5× Color All-in-one RT Buffer, 2 μ l of NC Buffer, 1 μ l of Oligo dT18 (20×), 1 μ l of Random Hexamer (20×), and then add ddH₂O (Nuclease) to up to 20 μ l:

Components	Volume (20 μl)
total RNA	1 µg
or poly(A) mRNA	200 ng
5× Color All-in-one RT Buffer	4 µl
NC Buffer	2 μΙ
gene-specific primer (2 µM)	1 μΙ
ddH ₂ O (Nuclease ⁻)	up to 20 μl

- 2. After the reaction is prepared according to steps 1A and 1B, mix thoroughly before the reverse transcription reaction (Important). It is recommended to adjust the volume of the pipette to about 18 μ l and gently pipette 10 \sim 15 times.
- 3. Incubate at 42°C for 15 min, then Incubate at 95°C for 30 s to inactivate DNase and reverse transcriptase.
- 4. The reverse transcription reaction product can be directly used in PCR applications, but the cDNA is recommended to dilute and mix thoroughly before use (The specific dilution factor depends on the abundance of gene expression. Generally, cDNA is diluted $5 \sim 10$ times). If the qPCR experiment is not performed immediately, it is recommended to store at -80°C for long-term storage. Avoid repeated freeze-thaw cycles.